



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: Jan 12, 2005 INITIALS: DEC OWNER: WSP
DISTRICT: 3 ADDRESS: _____
LOCATION: 7 M Shop COUNTY: _____
MAKE: DEC TYPE: port-wheel weigher ACCURACY CLASS: III
MODEL: MD-400 ID. NO.: 17135 CAPACITY: 20000 lb SECTIONS: 1
MINIMUM GRADUATIONS: 20 lb REG. ELEMENT: dial PLATFORM: 11" X 12"

LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
<u>Center</u>	<u>5000</u>	<u>5000</u>	<u>0</u>				
<u>"</u>	<u>10000</u>	<u>10000</u>	<u>0</u>				
<u>"</u>	<u>15000</u>	<u>14960</u>	<u>-40</u>				
<u>Center</u>	<u>20000</u>	<u>20000</u>	<u>0</u>				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as is shown on the attached "Certificate of Accuracy", No. 19577

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cronin Date: 2/2/2005

State of Washington
County of Thurston
Signed and sworn to (or affirmed) before
me on February 2, 2005 by
Michael G. Evans
(Signature of notary public)
Michael J. Evans
Printed Name of Notary
My appointment expires 12-27-07



3000-362-021 (R) (10/98)

DISTRIBUTION: WHITE - COURT GREEN - DISTRICT CANARY - PROPERTY MANAGEMENT PINK & GOLDENROD - SERVICE TRUCKS